

doi: [10.4103/0971-5916.91012](https://doi.org/10.4103/0971-5916.91012)

[Copyright](#) : © The Indian Journal of Medical Research

Framework for estimating tumour parameters using thermal imaging

V. Umadevi, S.V. Raghavan, and Sandeep Jaipurkar*

Network Systems Laboratory, Indian Institute of Technology Madras, India

**Vijaya Scans, Chennai, India*

Reprint requests Dr V. Umadevi, BSB 371, Network Systems Laboratory, Department of Computer Science & Engg., Indian Institute of Technology Madras, Chennai 600 036, India ;

Email: umadevi.svr@cs.iitm.ernet.in

Received March 30, 2010

This is an open-access article distributed under the terms of the Creative Commons Attribution-Noncommercial-Share Alike 3.0 Unported, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

- [Other Sections ▼](#)

Abstract

Background & objectives:

Non-invasive and non-ionizing medical imaging techniques are safe as these can be repeatedly used on an individual and are applicable across all age groups. Breast thermography is a non-invasive and non-ionizing medical imaging that can be potentially used in breast cancer detection and diagnosis. In this study, we used breast thermography to estimate the tumour contour from the breast skin surface temperature.

Methods:

We proposed a framework called infrared thermography based image construction (ITBIC) to estimate tumour parameters such as size and depth from cancerous breast skin surface temperature data. Markov Chain Monte Carlo method was used to enhance the accuracy of estimation in order to reflect clearly realistic situation.

Results:

We validated our method experimentally using Watermelon and Agar models. For the Watermelon experiment error in estimation of size and depth parameters was 1.5 and 3.8 per cent respectively. For the Agar model it was 0 and 8 per cent respectively. Further, thermal breast screening was done on female volunteers and compared it with the magnetic resonance imaging. The results were positive and encouraging.

Interpretation & conclusions:

ITBIC is computationally fast thermal imaging system and is perhaps affordable. Such a system will be useful for doctors or radiologists for breast cancer diagnosis.

Keywords: Non-invasive, non-ionizing, thermography, tumour parameters

- [Other Sections ▼](#)

Breast thermography started during 1956¹, but the results of Breast Cancer Detection and Demonstration Project (BCDDP) done between 1973 and 1981 by American Cancer Society and National Cancer Institute of United States were not promising due to lack of trained professionals and also primitive thermal imaging techniques². Recent advances in sensor technology and awareness of using non-invasive and non-ionizing medical imaging techniques has made breast thermography a choice for breast cancer diagnosis.

Inference of tumour parameters from surface temperature distribution of cancerous skin is an inverse heat transfer problem which is “ill-posed”³. Methods have been proposed to find parameters of embedded tumours using surface temperature distribution⁴⁻⁶. Thermal texture method⁴ finds depth of tumour, but its principle is hard to comprehend⁷. Other methods^{5,6} are based on artificial neural network and genetic algorithm. These methods are time consuming and error prone in the presence of the noisy data⁸.

More than 1.2 million people are diagnosed with breast cancer each year worldwide and over 500,000 die from this disease⁹. Predictions of breast cancer development for the next 10 years⁹ based on cases diagnosed during 2000-2002 world wide has been done¹⁰.

On January 29, 1982 United States Food and Drug Administration (FDA) has approved breast thermography for breast cancer detection and diagnosis¹¹. Earlier studies¹²⁻¹⁴ on breast thermography showed both positive and negative aspects of the technology.

Intensities of infra red (IR) radiations emitted by human body vary in accordance to the temperature of the body. The process of capturing, converting, and recording IR radiation emissions from human body is referred as thermal imaging or thermography which is used in medicine for vascular disorder detection, chronic pain detection, assessment of skin burn, dentistry, skin cancer detection, breast cancer detection and many others¹⁵⁻¹⁹. Biologically, metabolic rate of cancerous cells is comparatively higher than normal cells. Hence, tumours act as heat source increasing surface temperatures around the cancerous area which could be seen as a hot spot in a thermal image.

Here we propose the infrared thermography based image construction (ITBIC) framework to find parameters *i.e.* size and depth, of a tumour. We used Markov Chain Monte Carlo (MCMC) numerical method²⁰ to enhance the accuracy in parameter estimation. Results were obtained experimentally using synthetic tissue models (watermelon and agar) as well as direct clinical trials using volunteer subjects.

- [Other Sections ▼](#)

Material & Methods

Two different infrared cameras namely Ti40FT from M/s Fluke Corp., USA and Varioscans 3021 ST from Jenoptik Laser, Germany, were used for breast thermal scanning. Specifications of these two infrared cameras are given in [Table I](#). Fifty women of various age groups were screened for breast thermal scanning at Vijaya scans, Chennai, after informed consent was taken from the patients during May 2007 - March 2010. The study protocol was approved by the ethics committee of the Paterson Cancer Center, Chennai. For each breast thermal images were captured of front, left and right side views. The first 27 sets were screened using Ti40FT and second 23 sets were screened using Varioscans 3021 ST. Procedure followed for screening of first and second sets were different. First 27 sets were screened in a room which was air conditioned and with fluorescent bulb on. To normalize, subjects were made to sit in the examination room for 15 min with upper body part undressed. Then the thermal image of the breast part were captured. Second 23 sets were screened in a room which was not air conditioned and with fluorescent bulb off (*i.e.* images were captured in a dark room). For this set of subjects images were obtained immediately (*i.e.* without normalizing) with upper body part undressed. Subject were found to be more comfortable to thermographic procedure when carried out in a dark room.

Table I


Specifications of infrared camera

	Fluke Ti40FT	Varioscans 3021 ST
Spectral range (μm)	8 to 14	8 to 12
Detector type	Focal plane array, Vanadium oxide uncooled microbolometer	Mercury-Cadmium-Telluride (MCT) single element detector Stirling cooling
Thermal sensitivity (mK)	90	30
Image size (pixels)	160 x 120	360 x 240
Spatial resolution (mrad)	2.60	1.5
Field of view	23° horizontal x 17° vertical	(30x20)° to (5x3.3)° with electro optical

Direct estimation of tumour parameters from breast thermal image is a tedious task. We proposed a framework named ITBIC to estimate tumour parameters from surface temperature data ([Fig. 1](#)).

Fig. 1

Framework for tumour parameter estimation.



In the initial ITBIC framework^{21,22}, case of radiative heat transfer model was considered tissue (breast) as cube and tumour as a 2D model *i.e.*, a disc. Subsequently, breast tissue was considered as a 3D hemisphere model and tumour as a sphere model to improve the accuracy of ITBIC framework. Pennes bio-heat transfer equation²³ was used to model heat transfer process within the tissue. [Table II](#) gives the values of the tissue and tumour considered by our estimation

system for solving Pennes heat transfer equation. COMSOL²⁴ software was used for solving Pennes heat transfer equation.

Table II

Pennes bio-heat transfer equation parameters values

Parameter name	Symbol	Tissue	Tumour
Thermal conductivity (W/m)	k	0.42	0.42
Blood perfusion rate (1/s)	ω_b	18e ⁻⁶	9e ⁻⁶
Density (kg/m ³)	ρ_b	920	920
Specific heat of blood (J/kg.K)	c_b	3000	3000
Arterial blood temperature (K)	T _a	310	310
Metabolic heat generation rate (W/m ³)	Q _{met}	450	29000

ITBIC framework works in two stages, first it estimates actual size & location of tumour and second stage calculates depth.

Estimation of size and location: ITBIC frame work is supplied with input of cancerous surface temperature measurements, T_{meas} and then initial assumption for tumor size (r₀) and location (x₀ y₀ z₀) within the hemisphere (breast model). Using COMSOL software, Pennes bio-heat transfer equation is solved to generate surface temperature measurement, T_{gen}, for the initial assumed tumour size (r₀) and location (x₀ y₀ z₀). Then using Metropolis-Hastings algorithm, a MCMC method²⁰, next estimate of the tumour size (r_i) and location (x_i y_i z_i) is generated depending on the comparison result between T_{meas} and T_{gen}. This procedure is repeated in a loop until a stable estimation for tumor size I and location (x y z) is found.

Calculation of depth: Using the estimated values of size I and location (x y z) of the tumour inside the breast, depth of the tumour is calculated. Considering the center co-ordinates of the hemisphere as (x_c y_c z_c), estimated location co-ordinates of tumour inside the breast as (x y z) and coordinates on the border surface of the breast as (x_b y_b z_b), value of x_b is equivalent to x and value of y_b is equivalent to y. Depth of the tumor from the top surface of the breast to be calculated is D, which is as follows:

$$R^2 = (x_b - x_c)^2 + (y_b - y_c)^2 + (z_b - z_c)^2$$


$$z_b = \sqrt{R^2 - ((x_b - x_c)^2 + (y_b - y_c)^2)} + z_c$$

$$D = (z_b - z - r)$$

Study with phantom tissue: Framework proposed was validated first with synthetically designed phantom tissue and tumour materials and then clinical studies were carried out to validate. Laboratory experiments were conducted for the study of homogeneous and heterogeneous nature of tissue *i.e.* Agar model experiment to represent homogeneous and Watermelon experiment to represent heterogeneous nature. The experimental apparatus and set up are shown in [Fig. 3](#).

Fig. 3

Laboratory experimental setup with phantom tissue and tumour material.



Electric bulb was used with phantom tissues to mimic tumour. Diameter of the electric bulb was 1.28 cm. Phantom tissue with bulb inserted was kept on a thermocol sheet. Electric bulb was connected to power supply unit.

Agar model experiment: Agar powder was boiled in water around 30 min and poured into a plastic jar which contained 4.5 watt electric bulb. The solution was then allowed to cool down to room temperature. At the steady state of room temperature the size of cast model obtained was $10 \times 10 \times 9$ cm (length \times width \times height), which was shaped like a cube to mimic tissue.

Watermelon experiment: Watermelon pulp of size $10 \times 10 \times 10$ cm (length \times width \times height) was used. Pulp was shaped like a cube to mimic tissue. An electric bulb of 4.5 watts was introduced into the cube to mimic tumour.


- [Other Sections](#) ▼

Results

Example of an normal and abnormal breast thermal image obtained from women volunteers are shown in Figs [Figs2a2a](#) and [and2b2b](#) as captured by the Varioscan 3021 ST camera at a temperature range of 32-38°C.

Fig. 2

(a). Normal and (2b). abnormal breast thermogram



Temperature distribution between left and right parts of the breast was symmetrical for a normal case (45 yr old healthy woman) ([Fig. 2a](#)) (as the colour pattern is symmetrical). There will be rise in temperature of the breast skin surface area above the existence of cancerous mass.

Thermogram of an abnormal case (female aged 54 yr) ([Fig. 2b](#)) showed asymmetrical colour pattern. Her self breast examination showed lump in her left breast part, so she reported to clinic for further examination. MRI scan report of this woman confirmed having cancerous mass in left part of the breast. In the thermal image, left part of the breast showed the increased surface temperature when compared to right part of the breast. Further, biopsy test conducted on this patient also proved lump as malignant neoplasm.

With the experimental set up, the electric bulb was switched on for around 30 min and then thermal images were captured using the camera Fluke Ti40FT for Agar and watermelon

experiments (Figs (Figs4,4, ,5).5). Twenty five surface temperatures, from the top and from a side of agar model were measured. This measurement values were given as an input to ITBIC framework to estimate radius and depth of the tumour. Estimated parameter values against actual values are given in [Table III](#). The estimated values were very close to the actual values.

Fig. 4

Infrared images of agar model with surface temperature distribution.




Fig. 5

Watermelon model surface temperature distribution.




Table III

ITBIC framework result for Agar model and watermelon experiment

	Tumour size (Radius, cm)		Depth from top (cm)		Depth side
	Actual	Estimated	Actual	Estimated	Actual
Agar model Agar model	0.44	0.44	2.00	1.72	0.00
Watermelon experiment	0.44	0.44	0.00	0.00	0.00

Among 50 women, two had a focal area of raised temperature in left and right breasts respectively. The same subjects were scanned by MRI and also biopsy was carried out, both tests proved the area to be malignant neoplasm. The temperature difference between normal breast surface and the neoplastic zone was greater than 2°C.

Thermal image and MRI of the first case are shown in [Fig 6a](#) and [Fig 6b](#). She was 54 yr old and had menopause at the age of 44. In the upper part of the left breast (which is marked by a block) tumour was present. Temperature measurements of this skin surface were used as an input to our ITBIC framework to estimate tumour size and depth. Estimated tumour size and depth are given in [Table IV](#). Results of our framework correlated closely with the MRI findings.

Fig. 6

(a). Breast thermal scan image of *Case 1* showing the abnormality. (b) MRI scan image of *Case 1* showing the same abnormality.




Table IV


ITBIC framework results for case 1 and 2 breast thermogram

	Tumour size (radius, cm)		Depth from top (cm)	
	Actual	Estimated	Actual	Estimated
Case 1:	1.50	1.59	2.50	2.46
Case 2:	1.00	1.44	2.00	1.85

Thermal image and MRI of the second case shown in [Fig. 7a](#) and [Fig. 7b](#). She was around 37 yr old. In the lower part of the right breast (which is marked by a block) the tumour was present. Estimated tumour size and depth are given in [Table IV](#).

Fig. 7

(a) Breast thermal scan image of *Case 2* showing the abnormality. (b) MRI scan image of *Case 2* showing the same abnormality.



Of the 50 female subjects, 44 were normal, two had malignancy, three had fibrocystic disease and one was abscess case. With thermography, temperature difference between left and right parts of the breast in each of these cases were observed ([Table V](#)).

Table V

Observed breast temperature variation

Category	Collateral temperature (°C) difference
Normal	< 2
Fibrocystic disease	2 to 3
Malignant + Inflammatory	> 3

- [Other Sections ▼](#)

Discussion

Presently commonly used medical imaging techniques for breast cancer are mammography, MRI and ultrasound, each having their own advantages and limitations.

Our improved ITBIC framework surpassed all the previous vogue for methods in tumour parameter estimation^{5,6,25} in terms of accuracy and simulation time.

The results of ITBIC framework in comparison with conventional method *i.e.* MRI showed a promising relation in estimation of tumour parameters from breast thermograms. Further study of ITBIC framework on large data set need to be done to judge its performance.

In conclusion, a framework was developed to identify parameters of tumour in human female breasts from cancerous skin surface temperature. Information on the size and depth of a tumour may provide critical information to medical professionals for diagnosis. Tumour size and depth parameter values were estimated by ITBIC framework from surface temperature measurements of breast. Estimated values matched with actual values given by MRI scanning method. This shows practicability of using ITBIC framework by medical professionals to diagnose the breast tumour.

Currently, the procedure of breast thermography is not practiced or less in use. Progressively, the use of breast thermography for scanning of subjects may supercede the other scanning methods due to its advantages.

Acknowledgments

Authors thank Dr Sandeep Jaipurkar, Vijaya Scans, Chennai, for his inputs and support provided during the process of experimental clinical study. Authors also thank Dr S. Vijayaraghavan, Paterson Cancer Center, Chennai, and staff of Vijaya Scans, Chennai, for their support during the process of breast thermal imaging. Authors also thank Dr Ranganath R. Navalgund, Space Applications Centre, ISRO, Ahmedabad and Prof. V. Jagadeesh Kumar, Department of Electrical Engineering, IIT Madras, Chennai, for sharing their infrared cameras with us for our work.

- [Other Sections ▼](#)

References

1. Lawson R. Implications of surface temperatures in the diagnosis of breast cancer. *Can Med Assoc J.* 1956;75:309–10. [[PMC free article](#)] [[PubMed](#)]
2. Head JF, Elliott RL. Infrared imaging: making progress in fulfilling its medical promise. *IEEE Engineering Med Biol Magazine.* 2002;21:80–5.
3. Tikhonov AN. Washington D.C.: Halsted Press; 1977. Solution of Ill-posed problems.
4. Qi H, Kuruganti PT, Liu Z. Washington D.C.: 2002. Jul 7-12, Early detection of breast cancer using thermal texture maps. *IEEE Symposium on Biomedical Imaging: Macro to Nano.*
5. Paruch M, Majchrzak E. Identification of tumor region parameters using evolutionary algorithm and multiple reciprocity boundary element method. *Eng Appl Artif Intell.* 2007;20:647–55.
6. Mittal M, Pidaparti RM. Breast tumor simulation and parameters estimation using evolutionary algorithms. *Modell Simul Eng.* 2001;756436–41. Article I.
7. Shang Z, Jiang G. Fundamental theoretic research of thermal texture maps I - simulation and analysis of the relation between the depth of inner heat source and surface temperature distribution in isotropy tissue. *Conf Proc IEEE Eng Med Biol Soc.* 2004;7:5271–3. [[PubMed](#)]
8. Whitley D. Genetic algorithms in engineering and computer science. UK: John Wiley; 1995. Genetic algorithms and neural networks; pp. 2003–16.
9. Askoxylakis V, Thieke C, Pleger ST, Most P, Tanner J, Lindel K, et al. Long-term survival of cancer patients compared to heart failure and stroke: a systematic review. *BMC Cancer.* 2010;10:105. [[PMC free article](#)] [[PubMed](#)]
10. Ferlay J, Bray F, Pisani P, Parkin DM. Lyon, France: IARC Press; 2004. Cancer incidence, mortality and prevalence Worldwide, GLOBOCAN 2002, IARC Cancer Base, No. 5, version 2.0.
11. Amalu WC. A review of breast thermography. CA, USA: International Academy of Clinical Thermology; 2002.
12. Gutierrez-Delgado F, Vazquez-Luna J, Venegas-Hernandez L, Terrazas-Espitia S, Marcial-Toledo S, Guzman-Patraca C, et al. Feasibility of thermal infrared imaging screening for breast cancer in rural communities of Southern Mexico: The experience of the Centro de Estudios y Prevencion del Cancer (CEPREC) *J Clin Oncol.* 2009;27(Suppl):1521.
13. Ng EYK. A review of thermography as promising non-invasive detection modality for breast tumor. *Int J Thermal Sci.* 2009;48:849–59.
14. Williams KL, Phillips BH, Jones PA, Beaman SA, Fleming PJ. Thermography in screening for breast cancer. *J Epidemiol Commun Health.* 1990;44:112–3.
15. Bagavathiappan S, Saravanan T, Philip J, Jayakumar T, Raj B, Karunanithi R, et al. Investigation of peripheral vascular disorders using thermal imaging. *Br J Diabet Vascul Dis.* 2008;8:102–4.
16. Hildebrandt C, Raschner C, Ammer K. An overview of recent application of medical infrared thermography in sports medicine in Austria. *Sensors.* 2010;10:4700–15.
17. Ruminski J, Kaczmarek M, Renkielska A, Nowakowski A. Thermal parametric imaging in the evaluation of skin burn depth. *IEEE Trans Biomed Eng.* 2007;54:303–12. [[PubMed](#)]
18. Komoriyama M, Nomoto R, Tanaka R, Hosoya N, Gomi K, Iino F, et al. Application of thermography in dentistry--visualization of temperature distribution on oral tissues. *Dental Mater J.* 2003;22:436–43.

19. Buzug TM, Schumann S, Pfaffmann L, Reinhold U, Ruhlmann J. Functional infrared imaging for skin-cancer screening. *IEEE Eng Med Biol Soc.* 2006;1:2766–9.
20. Parthasarathya S, Balaji C. Estimation of parameters in multi-mode heat transfer problems using Bayesian inference - Effect of noise and a priori. *Int J Heat Mass Transfer.* 2008;51:2313–34.
21. Umadevi V, Raghavan SV. Proceedings of the 3rd International Conference on Bioinformatics and Biomedical Engineering. Beijing, China: 2009. Jun 11-13, Infrared thermography based image construction for biomedical applications.
22. Umadevi V, Suresh S, Raghavan SV. Proceedings of the 31st Annual International Conference of the IEEE Engineering in Medicine and Biology Society. Minneapolis, Minnesota, USA: 2009. Sep 2-6, Improved infrared thermography based image construction for biomedical applications using Markov Chain Monte Carlo Method.
23. González FJ. Thermal simulation of breast tumors. *I Rev Mex Fís.* 2007;53:323–6.
24. Burlington, MA: 2007. COMSOL Multiphysics, Version 3.4.
25. Li KY, Dong YG, Chen C, Zhang SP. The noninvasive reconstruction of 3D temperature field in a biological body by Monte Carlo Method. *J Neurocomputing.* 2008;72:128–33.

Articles from The Indian Journal of Medical Research are provided here courtesy of
Medknow Publications